


INSTRUMENT

4A

BEAM LINE

Fact Sheet




MAGNETISM REFLECTOMETER

SPECIFICATIONS

Source-sample distance	18.64 m
Sample-detector distance	0.5 – 6 m
Detector size	20 • 20 cm ²
Detector resolution	1.5 mm
Moderator	coupled supercritical hydrogen
Bandwidth	$\Delta\lambda = 3.1 \text{ \AA}$
Wavelength range	$1.8 \text{ \AA} < \lambda < 14.0 \text{ \AA}$
Q range	$0 \text{ \AA}^{-1} < Q < 7.0 \text{ \AA}^{-1}$
Minimum reflectivity	$10^{-9} - 10^{-10}$
Detector area	18 cm x 18 cm

determination for direct inversion of reflectivity data into real-space scattering-length density profiles.




RECENT SIGNIFICANT EVENTS

Instrument Construction

- Instrument construction is complete.
- The instrument safety review was conducted on June 1, 2006.
- Commissioning will start in July 2006.

Instrument Science

- Outfitting of a lab for neutron polarization equipment is under way.
- A workshop on high-magnetic field science was held at the National High Magnetic Field Laboratory.



Magnetism reflectometer

FOR MORE INFORMATION, CONTACT MAGNETISM REFLECTOMETER STAFF

Instrument Scientist: Frank Klose, klosefr@sns.gov, (865) 576-5389

Lead Engineer: Tim Chae, chaet@sns.gov, (865) 241-6740

Electrical Engineer: Andre Parizzi, parizziad@sns.gov, (865) 576-6219

Scientific Associate: Richard J. Goyette Jr., goyetterj@sns.gov, (865) 241-9991

www.sns.gov/users/instrument_systems/instruments/elastic/magnet.shtml



June 2006

05-03085A/arm